

Report on SFI Measurements

James Knowlton
Division Director
Standardization & Engineering Division
Cotton & Tobacco Program
USDA AMS
Memphis, Tennessee, USA



Values for SFI Calibration

- Unofficial values available upon request for SFI on valid calibration cottons produced since April 2008
(Except ELS calibration cottons)
- Requests have continued to increase each year
- Majority of calibration cottons go out without SFI values requested

SFI Results from CSITC RT

<u>RT</u>	<u>SFI SD</u>	<u>SFI CV</u>
2011-1	1.58	19.9
2011-2	1.89	19.5
2011-3	1.64	19.9
2011-4	2.09	20.7
Avg.	1.80	20.0
2012-1	1.80	16.7
2012-2	1.42	14.7
2012-3	1.06	11.5
2012-4	1.24	12.9
Avg.	1.38	14.0
2013-1	1.45	15.7
2013-2	1.35	11.6
2013-3	1.07	11.5
2013-4	1.15	11.3
Avg.	1.26	12.5

USDA HVI Check Test – Average CV's
(within + between instrument)

